

State Level Water Testing Laboratory

PWD(DWS), Government of Tripura
P.N. Complex,
Agartala - 799006
Mobile : 9402120089,
Website : www.dws.tripura.gov.in



TC-7350

TEST REPORT

Name & Address of the customer

Vice Principal, Henry Derozio
Academy, Water sample from Henry
Derozio Academy, Kunjaban,
Agartala, Tripura

Report No : SLWTL/2017/4894
Report Date : 03-09-2025
Sample ID : ARJ170FWOX
Date of Receipt of Sample : 01/09/2025, 11.30 am
Date(s) of Performance : 02/09/2025
Ref. No. & Date : No. F. 3(1)-HDA-HS/2025 dated 18/08/2025
ULR No : TC735025000011886F

We hereby certify that the following sample submitted to the laboratory has been analyzed with the following results :

1.	Description of Sample (As declared by customer)	: Drinking Water
2.	Volume of Sample Received	: 600 ml
3.	Sample Mark (if any, given by the customer)	: NIL
4.	Date of Sampling by the Customer	: 01/09/2025
5.	Source & Place of Sampling	: Water sample from Henry Derozio Academy, Kunjaban, Agartala
6.	Environmental conditions during sampling	: NA
7.	Sampling Plan & Procedures used	: NA

Report Prepared by

Signature

Name : Samir Acharjee

Designation : Chemist

For State Level Water Testing Laboratory

Signature

Name : Sujoy Paul

Designation : Technical Manager

Authorized Signatory

- The results relate only to the item(s) tested.
- This Test Report shall not be reproduced except in full, without the permission of State Level Water Testing Laboratory, PWD(DWS)
- The reserved part of sample(s) shall be retained for 15 days from the date of Completion of Test.

Contact Person : Er. Rajib Majumder,
Quality Manager, State Level Water Testing Laboratory
Mobile : 9402120089, Email : qmgtripura@gmail.com, Website : www.dws.tripura.gov.in

State Level Water Testing Laboratory

PWD(DWS), Government of Tripura
P.N. Complex,
Agartala - 799006
Mobile : 9402120089,
Website : www.dws.tripura.gov.in



TC-7350

Report No : SLWTL/2017/4894
Sample ID : ARJ170FWOX

Analysis of Result

SN	Characteristic	Acceptable Limit (as per BIS 10500:2012)	Permissible Limit (as per BIS 10500:2012)	Method of Test	Results
1	Turbidity in N.T.U	1	5	APHA 24th Edition 2130 B	0.65
2	pH value at 25°C	6.5-8.5	No Relaxation	APHA 24th Edition 4500-H+ B	7.22
3	Iron (as Fe) in mg/l	1.0	No Relaxation	APHA 24th Edition 3500-Fe B	0.25
4	Total Dissolved Solids in mg/l	500	2000	APHA 24th Edition 2540 C	69.6

BDL - Below Detection Level for Fluoride - <0.2 mg/l, Arsenic - <0.009 mg/l, Iron - <0.1 mg/l, Magnesium - <2 mg/l, Manganese - <0.01 mg/l

Report Prepared by

For State Level Water Testing Laboratory

Signature 
Name : Samir Acharjee
Designation : Chemist

Signature 
Name : Sujoy Paul
Designation : Technical Manager
Authorized Signatory

-: End Of Report :-

Contact Person : E2. Rajib Majumder
Quality Manager, State Level Water Testing Laboratory,
Mobile : 9402120089, Email : qm@tripura@gmail.com, Website : www.dws.tripura.gov.in